



Ref. Certif. No.

SE-51251IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE****CERTIFICAT D'ESSAI OC**Product
Produit

Laser device

Name and address of the applicant
Nom et adresse du demandeurDisMark GmbH
Rellikonstrasse 7,
CH-8124 Maur, SwitzerlandName and address of the manufacturer
Nom et adresse du fabricant

Same as above

Name and address of the factory
Nom et adresse de l'usine

Same as above

Ratings and principal characteristics
Valeurs nominales et caractéristiques principalesInternally powered (2x1.5V LR3 AAA Alkaline Batteries), applied
part type B, laser product class 3R, 5mW, 650nm.Trademark (if any)
Marque de fabrique (si elle existe)

-

Model / Type Ref.
Ref. De type

MedicLaser

Additional information (if necessary)
Les informations complémentaires (si nécessaire)

See page 2

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la


IEC 60601-1:1988 and A1+A2

As shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

07KFM007721

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00**Intertek ETL SEMKO**

Date: 13 August 2007

Signature: 

Additional information (if necessary)
Information complémentaire (si nécessaire)

Model MedicLaser is part of a complete delivery package for treatment of inner ear diseases, which is sold as Model EarLaser. Model EarLaser consists of Model MedicLaser (class 3R Laser Product), and accessory Model EarTool.

Date: 13 August 2007

Signature:

